

Bruksanvisning – Catell Active Shoulder

Catell rekommenderar att produkten provas ut och anpassas av legitimerad eller övrigt utbildad sjukvårdskunnig personal. Brukare med diabetes, känslig hud, kända kontaktallergier, nedsatt känsel eller dålig cirkulation ska vara extra observant för tryck, skav eller ödem under användning. Kontrollera alltid materialinnehåll nedan. Alla allvariga incidenter i samband med användning av denna produkt ska anmälas till tillverkaren och till behörig myndighet i den medlemsstat där användaren och/eller patienten är etablerad. Om frågor kring produkten uppstår kontakta din ordinator, utprovare eller Catell.

Användningsområde

Ortosen stödjer och avlastar brukarens arm och skuldra efter skada eller operation.

Applicering

1. Öppna armdelens och bröstbandets kardborrband. Trä sedan armen genom ortosen, applicera på axeln. Bild 1
2. Fäst och justera bröstbandet genom att trä det genom plastöglan på framsidan. Bild 2
3. Justera armdelens kardborrstängning. Bild 3
4. Säkerställ att samtliga kardborrband är ordentligt stängda.



Ortosen kan behöva justeras med jämna mellanrum beroende på vilka situationer den används i. Det är viktigt att ortosen sitter bekvämt, utan att klämma, skava eller orsaka stas.

Material

Polyester, polyamid, elastan. Produkten är latexfri.

Tvättråd

Slut kardborrbanden. Följ tvättråden nedan. Använd gärna tvättpåse. Låt ortosen droptorka.



Måttabell

Strl	*Mått	Strl	*Mått	Strl	*Mått
XS	25 - 27 cm	M	29 - 32 cm	XL	35 - 38 cm
S	27 - 29 cm	L	32 - 35 cm		

*Mått tas runt överarmen och ska ses som en vägledning.

Produkten är bilateral.

Denna produkt är endast avsedd att användas på en patient.

Återvinning

Separera plast och textil. Sortera enligt lokala föreskrifter.

Denna produkt överensstämmer med de krav som ställs i den Medicintekniska förordningen (EU) 2017/745, klass I produkter.



Manufacturer Catell AB

Örtökron 23, SE-138 40 Älta, Sweden | Phone: +46 8 727 95 30 | info@catell.se | www.catell.se

240411:5

Instruction for use – Catell Active Shoulder



Catell recommend that the product is fitted and adapted by a licensed or otherwise qualified health professionals. Users with diabetes, sensitive skin, known contact allergies, impaired sensibility or poor circulation must be extra observant and check for pinching, chafing or edema when using the brace. Always check material content below. Any serious incidents occurring relates to the device should be reported to the manufacturer and to the competent authority of the Member State in which the user and/or patient is resident. Contact your prescriber or Catell if questions about the product occur.

Area of use

The brace supports the users arm and shoulder after injury or surgery.

Application

1. Open the hook and loop straps on the arm part and on the front part. Place the brace on the shoulder by pulling it up the arm. Picture 1
2. Attach and adjust the chest strap through the front parts plastic buckle. Picture 2
3. Adjust the hook and loop straps on the arm part. Picture 3
4. Make sure that all straps are properly secured.



When necessary adjust the brace. It is important that the brace fits comfortably without pinching, chafing or affecting blood circulation.

Material

Polyester, polyamide, elastane. Free from latex.

Washing instructions

Close the hook and loop straps. Follow the washing instructions below. Use a washing pouch. Let the brace drip dry.



Size

Size	*Dimensions	Size	*Dimensions	Size	*Dimensions
XS	25 - 27 cm	M	29 - 32 cm	XL	35 - 38 cm
S	27 - 29 cm	L	32 - 35 cm		

*Measure upper arm circumference. Measurements should be seen as guidelines.
The product is bilateral.
Single patient use.

Recycling

Separate plastic and textile. Sort according to local regulations.

This device conforms to the requirements of the Medical Device Regulation (EU) 2017/745 product class I.



Manufacturer Catell AB

Orkroken 23, SE-138 40 Älta, Sweden | Phone: +46 8 727 95 30 | info@catell.se | www.catell.se

240411:5