

# Bruksanvisning – Catell Armsling Double

Catell rekommenderar att produkten provas ut och anpassas av legitimerad eller övrigt utbildad sjukvårdskunnig personal. Brukare med diabetes, känslig hud, kända kontaktallergier, nedsatt känsel eller dålig cirkulation ska vara extra observant för tryck, skav eller ödem under användning. Kontrollera alltid materialinnehåll nedan. Alla allvarliga incidenter i samband med användning av denna produkt ska anmälas till tillverkaren och till behörig myndighet i den medlemsstat där användaren och/eller patienten är etablerad. Om frågor kring produkten uppstår kontakta din ordinarie läkare, utprovare eller Catell.

## Användningsområde

Ortosen stödjer och avlastar brukarens arm och skuldra efter skada eller operation.

## Applicering

1. Justera först armpåsen med dess breda kardborre så att den passar armen.
2. Öppna upp de långa kardborrebanden och placera underarmen i armpåsen. Bild 1
3. För de långa banden över axlarna, korsade dem över ryggen och för dem runt underarmen och fäst fram på armpåsen. Var noga med att justera in rätt längd för optimalt stöd. Bild 2a, 2b
4. Vid behov av extra fixering av armen mot bålen, använd det långa breda extrabandet. Bild 3
5. Säkerställ att samtliga kardborreband är ordentligt stängda.



Ortosen kan behöva justeras med jämna mellanrum beroende på vilka situationer den används i. Det är viktigt att ortosen sitter bekvämt, utan att klämma, skava eller orsaka stas.

## Material

Nylon, polyester. Produkten är latexfri.

## Tvättråd

Slut kardborrebanden. Följ tvättråden nedan. Använd gärna tvättpåse. Låt ortosen dropporka.



## Måttabell

Strl	*Mått	Strl	*Mått	Strl	*Mått
S	28 - 33 cm	M	33 - 38 cm	L	38 - 43 cm

\*Mät underarmens längd. Mått ska ses som en vägledning.

Produkten är bilateral.

Denna produkt är endast avsedd att användas på en patient.

## Återvinning

Separera plast och textil. Sortera enligt lokala föreskrifter.

Denna produkt överensstämmer med de krav som ställs i den Medicintekniska förordningen (EU) 2017/745, klass I produkter.

# Instruction for use – Catell Armsling Double



Catell recommend that the product is fitted and adapted by a licensed or otherwise qualified health professionals. Users with diabetes, sensitive skin, known contact allergies, impaired sensibility or poor circulation must be extra observant and check for pinching, chafing or edema when using the brace. Always check material content below. Any serious incidents occurring relates to the device should be reported to the manufacturer and to the competent authority of the Member State in which the user and/or patient is resident. Contact your prescriber or Catell if questions about the product occur.

## Area of use

The brace supports the users arm and shoulder after injury or surgery.

## Application

1. First adjust the arm pouch with its wide hook closure so that it fits the intended arm.
2. Open the hook and loop closures and place the forearm in the pouch. Picture 1
3. Pull the long straps over the shoulders, cross them behind your back, place around the forearm and attach the hook closure on the arm pouch. Thoroughly adjust the correct strap length for optimal support. Picture 2a, 2b
4. In need of additional fixation of the arm to the body, use the long wide additional strap. Picture 3
5. Make sure that all straps are properly secured.



When necessary adjust the brace. It is important that the brace fits comfortably without pinching, chafing or affecting blood circulation.

## Material

Nylon, polyester. Free from latex.

## Washing instructions

Close the hook and loop straps. Follow the washing instructions below. Use a washing pouch. Let the brace drip dry.



## Size

Size	*Dimensions	Size	*Dimensions	Size	*Dimensions
S	28 - 33 cm	M	33 - 38 cm	L	38 - 43 cm

\*Measure length of forearm. Measurements should be seen as guidelines. The product is bilateral. Single patient use.

## Recycling

Separate plastic and textile. Sort according to local regulations.

This device conforms to the requirements of the Medical Device Regulation (EU) 2017/745 product class I.



Manufacturer Catell AB  
Örkroken 23, SE-138 40 Älta, Sweden | Phone: +46 8 727 95 30 | info@catell.se | www.catell.se

240411:5