

Catell rekommenderar att produkten provas ut och anpassas av legitimerad eller övrigt utbildad sjukvårdskunnig personal. Brukare med diabetes, känslig hud, kända kontaktallergier, nedsatt känsel eller dålig cirkulation ska vara extra observant för tryck, skav eller ödem under användning. Kontrollera alltid materialinnehåll nedan. Alla allvarliga incidenter i samband med användning av denna produkt ska anmälas till tillverkaren och till behörig myndighet i den medlemsstat där användaren och/eller patienten är etablerad. Om frågor kring produkten uppstår kontakta din ordinator, utprovare eller Catell.

Användningsområde

Ortosen stödjer, avlastar och stabiliserar brukarens arm och skuldra efter skada eller operation.

Applicering

1. Vid behov, dra först på den medföljande strumpan på överarmen. Bild 1
2. Dra på ortosen på armen försiktigt, nerifrån och upp, placera på axeln. Bild 2
3. Fäst det långa övre bandet cirkulärt runt bröstkorgen. Bild 3
4. Justera armdelens stängningar och dess band så att ortosen sitter behagligt och stadigt. Bild 4
5. Säkerställ att samtliga kardborrband är ordentligt stängda.



Vid behov kan plasten justeras för bättre passform. Justera kardborrbanden vid behov. Det är viktigt att ortosen sitter bekvämt, utan att klämma, skava eller orsaka stas.

Material

PE-plast, polyester, nylon. Produkten är latexfri.

Tvättråd

Avlägsna textilpolstringen och slut kardborrbanden. Lägg dessa i en tvättpåse och följ tvättråden nedan. Plastdelen kan rengöras med tvål och vatten, ej maskintvätt. Låt ortosen dropportorka.



Måttabell

Strl	Cirkulärt mått överarm
One size, bilateral	24-40 cm

Denna produkt är endast avsedd att användas på en patient.

Återvinning

Separera plast och textil. Sortera enligt lokala föreskrifter.

Denna produkt överensstämmer med de krav som ställs i den Medicintekniska förordningen (EU) 2017/745, klass I produkter.



Manufacturer Catell AB

Örkroken 23, SE-138 40 Älta, Sweden | Phone: +46 8 727 95 30 | info@catell.se | www.catell.se

Instruction for use – Catell Humerus Lock



Catell recommend that the product is fitted and adapted by a licensed or otherwise qualified health professionals. Users with diabetes, sensitive skin, known contact allergies, impaired sensibility or poor circulation must be extra observant and check for pinching, chafing or edema when using the brace. Always check material content below. Any serious incidents occurring relates to the device should be reported to the manufacturer and to the competent authority of the Member State in which the user and/or patient is resident. Contact your prescriber or Catell if questions about the product occur.

Area of use

The brace supports and stabilizes the users arm and shoulder after injury or surgery.

Application

1. If desired, use the additional sleeve between the arm and the brace. Picture 1
2. Safely slide the brace on the arm and place on the shoulder. Picture 2
3. Attach the long upper strap around the chest. Picture 3
4. Adjust the arm parts straps and closures. Ensure a comfortably and tight fit. Picture 4
5. Make sure that all straps are properly secured.



If desired, the plastic parts can be adjusted. When necessary adjust the brace. It is important that the brace fits comfortably without pinching, chafing or affecting blood circulation.

Material

PE-plastic, polyester, nylon. Free from latex.

Washing instructions

Detach textile padding parts, close hook and loop straps. Put them in a washing pouch and follow the washing instructions below. The plastic part can be washed with soap and water, no machine wash. Let the brace drip dry.



Size

Size	Upper arm circumference
One size, bilateral	24-40 cm

Single patient use.

Recycling

Separate plastic and textile. Sort according to local regulations.

This device conforms to the requirements of the Medical Device Regulation (EU) 2017/745 product class I.



Manufacturer Catell AB

Örkroken 23, SE-138 40 Älta, Sweden | Phone: +46 8 727 95 30 | info@catell.se | www.catell.se

240523:6