

Catell rekommenderar att produkten provas ut och anpassas av legitimerad eller övrigt utbildad sjukvårdskunnig personal. Brukare med diabetes, känslig hud, kända kontaktallergier, nedsatt känsel eller dålig cirkulation ska vara extra observant för tryck, skav eller ödem under användning. Kontrollera alltid materialinnehåll nedan. Alla allvarliga incidenter i samband med användning av denna produkt ska anmälas till tillverkaren och till behörig myndighet i den medlemsstat där användaren och/eller patienten är etablerad. Om frågor kring produkten uppstår kontakta din ordinator, utprovare eller Catell.

Användningsområde

Ortosen stödjer, avlastar och stabiliserar brukarens CMC1- och MP-led efter skada, operation, kroniska smärttillstånd, ledförslitning eller överbelastningsbesvär. Ortosen kan även användas i förebyggande syfte vid tunga eller monotona rörelsemoment.

Applicering

1. Anpassa aluminiumramen i ortosen efter brukarens tumme och till önskad position. Bild 1
2. Öppna kardborrbanden och stäng tumgreppet. Placera handen i ortosen enligt bild 2.
3. Stäng kardborrbandet så att ortosen sitter behagligt. Justera tumgreppet. Bild 3



Klipp av överflödigt kardborrband för ett renare och snyggare utseende. Justera kardborrbanden vid behov. Det är viktigt att ortosen sitter bekvämt, utan att klämma, skava eller orsaka stas.

Material

Nylon, polyester, foam, aluminiumram. Produkten är latexfri.

Tvättråd

Slut kardborrbanden. Följ tvättråden nedan. Separat tvätt rekommenderas. Använd gärna tvättpåse. Låt ortosen droptorka.



Måttabell

Strl	*Mått
S/M	14 - 17 cm
M/L	17 - 20 cm

*Mått tas runt handleden och ska ses som en vägledning.

Produkten är bilateral. Denna produkt är endast avsedd att användas på en patient.

Återvinning

Sortera enligt lokala föreskrifter.

Denna produkt överensstämmer med de krav som ställs i den Medicintekniska förordningen (EU) 2017/745, klass I produkter.



Manufacturer Catell AB

Örskroken 23, SE-138 40 Älta, Sweden | Phone: +46 8 727 95 30 | info@catell.se | www.catell.se

Instruction for use – ThumbX

Catell recommend that the product is fitted and adapted by a licensed or otherwise qualified health professionals. Users with diabetes, sensitive skin, known contact allergies, impaired sensibility or poor circulation must be extra observant and check for pinching, chafing or edema when using the brace. Always check material content below. Any serious incidents occurring relates to the device should be reported to the manufacturer and to the competent authority of the Member State in which the user and/or patient is resident. Contact your prescriber or Catell if questions about the product occur.

Area of use

The brace support and stabilizes the users CMC1- and MP-joint of the thumb after injury, surgery, chronic pain, osteoarthritis or overuse problems. The brace can also be used for preventive purposes in heavy or repetitive movements.

Application

1. Adjust the aluminiumbar in the brace to the users thumb and desired position. Picture 1
2. Open all hook and loop straps, close the thumb strap. Place your hand in the brace as illustrated in picture 2.
3. Close all hook and loop straps so that the brace fits comfortably. Adjust the thumb grip. Picture 3



Cut off excess hook and loop straps for a cleaner and neater appearance. Adjust the hook and loop straps when necessary. It is important that the brace fits comfortably without pinching, chafing or affecting blood circulation.

Material

Nylon, polyester, foam, aluminiumbar. Free from latex.

Washing instructions

Close the hook and loop straps. Follow the washing instructions below. Separate wash is recommended. Use a washing pouch. Let the brace drip dry.



Size

Size	*Dimensions
S/M	14 - 17 cm
M/L	17 - 20 cm

*Measure wrist circumference. Measurements should be seen as guidelines.
Bilateral. Single patient use.

Recycling

Sort according to local regulations.

This device conforms to the requirements of the Medical Device Regulation (EU) 2017/745 product class I.

